



PROPOSED DOCUMENT

(revision of GHTF/SG1/N29:2005)

Title: Definition of the Term 'Medical Device'

Authoring Group: Study Group 1

Endorsed by: The Global Harmonization Task Force

Date: March 15th, 2011

Table of Contents

1.0	Introduction.....	4
2.0	Rationale, Purpose and Scope.....	5
2.1	Rationale	5
2.2	Purpose.....	5
2.3	Scope.....	5
3.0	References.....	5
4.0	Definitions.....	5
5.0	Definition of the terms 'medical device' and 'IVD medical device'.....	5
5.1	Medical device	5
5.2	IVD medical device	6

Preface

The document herein was produced by the Global Harmonization Task Force, a voluntary group of representatives from medical device regulatory authorities and the regulated industry. The document is intended to provide non-binding guidance for use in the regulation of medical devices, and has been subject to consultation throughout its development

There are no restrictions on the reproduction, distribution, translation or use of this document. However, incorporation of this document, in part or in whole, into any other document does not convey or represent an endorsement of any kind by the Global Harmonization Task Force.

1.0 Introduction

The objective of the Global Harmonization Task Force (GHTF) is to encourage convergence at the global level in the evolution of regulatory systems for medical devices in order to facilitate trade whilst preserving the right of participating members to address the protection of public health by those regulatory means considered the most suitable.

The primary way in which the Global Harmonization Task Force (GHTF) achieves its goals is through the production of harmonized guidance documents suitable for implementation or adoption by member Regulatory Authorities, as appropriate taking into account their existing legal framework, or by nations with developing regulatory programmes. Eliminating differences between jurisdictions decreases the cost of gaining regulatory compliance and allows patients earlier access to new technologies and treatments.

This guidance document is one of a series that together describe a global regulatory model for medical devices. It provides a definition of a term that is used in all GHTF publications. The GHTF first published guidance on this subject in 2005 entitled *GHTF/SG1/N29:2005 Definition of the Term 'Medical Device'*.

This document supersedes that previous version. It has been changed to:

- provide definitions for the terms 'IVD medical device' and 'accessory to a medical device'; and
- clarify the Notes to the definitions.

This document is intended for use by Regulatory Authorities, Conformity Assessment Bodies and industry, and will provide benefits in establishing, in a consistent way, an economic and effective approach to the control of medical devices in the interest of public health.

Regulatory Authorities that are developing regulations or amending existing ones are encouraged to consider the adoption of this guidance and the principles it embodies, as this will help to reduce the diversity of schemes worldwide and facilitate the process of harmonization.

This guidance document has been prepared by Study Group 1 of the Global Harmonization Task Force (GHTF). Comments or questions about it should be directed to the Chair of GHTF Study Group 1 whose contact details are available on the GHTF website¹.

¹ www.gh tf.org

2.0 Rationale, Purpose and Scope

2.1 Rationale

The development of consistent, harmonized definitions for a medical device, and an In Vitro Diagnostic medical device, that could be used within a global regulatory model would offer significant benefits to the manufacturer, user, patient or consumer, and to Regulatory Authorities and support global convergence of regulatory systems. Eliminating differences between jurisdictions decreases the cost of gaining regulatory compliance and allows patients earlier access to new technologies and treatments.

2.2 Purpose

This document is intended to provide harmonized definitions of the terms 'medical device' and 'In Vitro Diagnostic medical device'. These terms appear in guidance documents published by the Global Harmonization Task Force. Adopting the definitions from this document will allow a Regulatory Authority to identify the products subject to medical device regulatory controls.

This document is intended to serve as guidance for Regulatory Authorities, Conformity Assessment Bodies and the regulated Industry.

2.3 Scope

This document applies to products that have a medical purpose, including those used for the in vitro examination of specimens derived from the human body.

3.0 References

Not required for this document.

4.0 Definitions

Accessory to a medical device: means an article intended specifically by its manufacturer to be used together with that medical device to enable that device to be used in accordance with its intended use.

5.0 Definition of the terms 'medical device' and 'IVD medical device'

5.1 Medical device

Medical device means any instrument, apparatus, implement, machine, appliance, implant, in vitro reagent, software, material or other similar or related article:

a) intended by the manufacturer to be used, alone or in combination, for human beings for one or more of the specific purpose(s) of:

- diagnosis, prevention, monitoring, treatment or alleviation of disease,
- diagnosis, monitoring, treatment, alleviation of or compensation for an injury,

- investigation, replacement, modification, or support of the anatomy or of a physiological process,
- supporting or sustaining life,
- control of conception,
- disinfection of medical devices,
- providing information for medical or diagnostic purposes by means of in vitro examination of specimens derived from the human body;

and

- b) which does not achieve its primary intended action in or on the human body by pharmacological, immunological or metabolic means, but which may be assisted in its intended function by such means.

Note 1: Products which may be considered to be medical devices in some jurisdictions but for which there is not yet a harmonized approach, include:

- disinfection substances,
- aids for persons with disabilities,
- accessories to a medical device (see Note 2),
- components of a medical device,
- devices incorporating animal and/or human tissues.

Note 2: Some jurisdictions include accessories to a medical device within the definition of a medical device. Other jurisdictions do not adopt this approach but still subject an accessory to the regulatory controls (e.g. classification, conformity assessment, quality management system requirements etc.) that apply to medical devices.

5.2 IVD medical device

IVD medical device means a medical device, whether used alone or in combination, intended by the manufacturer for the in-vitro examination of specimens derived from the human body solely or principally to provide information for diagnostic, monitoring or compatibility purposes. IVD medical devices include reagents, calibrators, control materials, specimen receptacles, software, and related instruments or apparatus or other articles.

Note 3: In some jurisdictions, certain IVD medical devices may be covered by other regulations.