



IMDRF International Medical Device
Regulators Forum



25 YEARS AND BEYOND

30th IMDRF 2026

Day 2 IMDRF Stakeholder Forum | 15 September 2026

Restricted - مقيد

Regulatory Updates (Saudi Food & Drug Authority)

Abdullah Alghuraibi

Executive Director of Medical Devices Evaluation

Saudi Food and Drug Authority



IMDRF International Medical Device
Regulators Forum



HSA
Health Sciences Authority

25

Restricted
مقيّد
Saudi Food & Drug Authority




الهيئة العامة للغذاء والدواء
Saudi Food & Drug Authority

Contents

- ❑ **SFDA MD Regulatory Overview.**
- ❑ **Medical Device Marketing Authorization Requirements (MDMA).**
- ❑ **SFDA Regulatory Reliance Approach for Medical Devices.**
- ❑ **SFDA Published Regulatory Documents in (2026).**
- ❑ **Guidance Under Development.**
- ❑ **Recent Updates:UDI Compliance, Post-market Surveillance, Clinical Trials of Medical Devices, MD Marketing Authorization (MDMA).**



SFDA Medical Devices Regulation Framework



- ✓ **Support innovation** and medical devices technology development.
- ✓ **Enhance the Kingdom's leading role** internationally in the medical devices field.
- ✓ **Protect the public health & patient's safety** in KSA.
- ✓ **Support investments & encourages manufactures** to launch branches in the Kingdom.
- ✓ **Effective economic impact** for the Saudi market.



Medical Device Marketing Authorization Requirements (MDMA)

- ➡ Device Description and Specification, Including Variants and Accessories
- ➡ Design and Manufacturing Information
- ➡ Benefit-Risk Analysis and Risk Management
- ➡ Post Market Surveillance Plan

- ➡ Information to be Supplied by the Manufacturer
- ➡ Essential Principles of Safety and Performance
- ➡ Product Verification and Validation
- ➡ Periodic Safety Update Report (PSUR) and Post-Market Surveillance Report



SFDA Regulatory Reliance Approach for Medical Devices

Applicable to:

- Manufacturers
- Authorized Representatives (ARs)

Purpose:

Obtain SFDA Medical Device Marketing Authorization (MDMA) through the Reliance Pathway.

Recognized Regulatory Authorities (RRAs):

Country / Jurisdiction	Regulatory Authority
Australia	Therapeutic Goods Administration (TGA)
Canada	Health Canada
European Union (EU)	National competent authorities – European Economic Area (EEA)
Japan	Pharmaceuticals and Medical Devices Agency (PMDA)
United Kingdom (UK)	Medicines and Healthcare products Regulatory Agency (MHRA)
United States of America (USA)	Food and Drug Administration (FDA)

Note: The list of RRA may be updated periodically by the SFDA to include additional jurisdictions, including those of the IMDRF Management Committee (MC) members.



General Eligibility Criteria & Medical Device Exclusions

✓ ELIGIBLE IF...

- ✓ Valid RRA marketing authorization
- 📄 Device identical to RRA-authorized version
- 📋 Compliant with Saudi national requirements

✗ NOT ELIGIBLE IF...

- 👤 Based on self-declaration only
- ⚖️ Relies on predicate equivalence
- 👤 Custom-made for individual patients
- 🚨 Active Post-Market safety concerns or recalls




Finalizing requirements before submission



How to Apply

Apply through the GHAD Portal



Overview

2.1

Does the applied device and their Accessories have valid approval from other Country or Jurisdiction?:

Yes

2.1.1

Country and Approval Evidence

+ ADD NEW EVIDENCE

Enable Reliance Pathway

This selection formally declares the applicant's intent to utilize an RRA marketing authorization as supporting evidence for the SFDA's regulatory assessment.



IMDRF International Medical Device Regulators Forum




25

Restricted **مقيّد**
Ministry of Health
Saudi Food & Drug Authority



Reliance Form Submission



Review Times for Marketing Authorization

Standard Review

35–60

Working Days

Depending on:

- Product type
- Risk classification
- Complexity

Risk Classification & Lead Time

Risk Class	Lead Time
Class A	35 – 60 Working Days
Class B	35 – 60 Working Days
Class C	35 – 60 Working Days
Class D	35 – 60 Working Days

Reliance Pathway

Within the **Reliance Pathway**, review time may be reduced to 35 working days by relying on prior RRA approval instead of a full independent assessment, subject to eligibility and complete documentation



Reliance on Post- Market Surveillance

The SFDA applies reliance principles across several key post-market surveillance activities, including:

Definitions & Terminology	Adverse Event Coding Systems	Adverse event reporting requirements and timelines:	Safety Alert Detection & Information Sharing
- SFDA adopts and utilizes internationally recognized IMDRF definitions and terminology. - Supports consistency in regulatory communication and safety information exchange	- SFDA utilizes IMDRF adverse event terminology and coding systems. - Supports standardized categorization, identification of potential causes, and analysis of adverse event trends.	- Requirements and timelines are implemented using a risk-based approach. - Consider the nature and severity of the event and potential impact on patients, users, and public health.	- SFDA monitors safety information and regulatory actions, including recalls , field safety corrective actions, and other relevant safety communications. - Supports signal detection, risk prioritization, and assessment of further regulatory action in the Saudi market.



SFDA Published Regulatory Documents in (2026)

#	Document ID &Version No.	Document Title
1	MDS-REQ 3 (Ver. 3)	Requirements for Safe Use of Medical Devices Inside Healthcare Facilities
2	MDS-G028 (Ver. 1)	Guidance on Bundling Criteria for Medical Devices within a Single MDMA Application
3	MDS-G029 (Ver. 1)	Guidance on Risk Communication of Medical Devices



SFDA UnderDevelopment Regulatory Documents in (2026)


IN PROGRESS

#	Document ID &Version No.	Document Title
1	MDS-REQ 1 (Ver. 7)	Requirements for Medical Devices Marketing Authorization
2	MDS-REQ 4 (Ver. 2)	Requirements for the Import and Clearance of Medical Imaging Materials and Particle Accelerators Used in Radioisotopes Formation for Medical Applications
3	MDS-REQ 5 (Ver. 7)	Requirements on Importation and Shipments Clearance of Medical Devices and Supplies
4	MDS-REQ 9 (Ver. 3)	Requirements for Licensing of Medical Devices Establishments
5	MDS-REQ 11 (Ver. 3)	Requirements for Post-Market Surveillance of Medical Devices
6	MDS-G012 (Ver. 2)	Guidance on MDMA –Significant and Non-Significant Changes
7	MDS-G015 (Ver. 3)	Guidance on SFDA Requirements for Quality Assurance Programs for Radiation Emitting and Imaging Devices
8	MDS-G020 (Ver. 4)	Guidance on SFDA Recognized Standards Lists
9	New Document	Guidance on Medical Devices Labeling
10	New Document	Guidance on Biotechnology Based Medical Devices



SFDA MD-UDI Updates

➤ UDI Compliance



MDS – REQ 7

Requirements for Unique Device Identification (UDI)
for Medical Devices

of Devices

+ 500,000

of Manufacturers

+1950

Post-market Surveillance **Updates**


Safety Alerts

Safety Alerts	230 safety alerts affected the Saudi market out of 1,028 globally detected	
Action Types	Correction 545	Removal 479
Medical Devices Affected	2,941,538	

Adverse Events & Complaints

Received AE & Complaints reports	1279		
Type of Reports	Healthcare providers 322	Manufacturers & companies 807	Public 150

Officers of Healthcare providers: 1,724 Officers




Clinical Trials of Medical Devices Updates

During 2026, a total of 47 applications for medical device clinical trials were received & thoroughly evaluated

Since 2015, a total of 367 medical device clinical trial applications were submitted and reviewed, as shown in the table

Device Category




Marketing Authorization Statistics MDMA Updates


Statistics in 2026

# Request	7877
# Products	41637

Classification	# Products
A	16006
B	11875
C	10032
D	3724



Thank You!



**To access SFDA-MD
Regulations and
Requirements**



IMDRF International Medical Device
Regulators Forum



HSA
Health Sciences Authority

25

Restricted

