



IMDRF International Medical
Device Regulators Forum

Update on EU regulatory developments

Erik Hansson
European Commission

IMDRF – 15
19-21 March 2019
Moscow, Russia



The EU single market for medical devices



1. EU



2. EFTA/EEA:
Norway, Liechtenstein, Iceland



3. Turkey



4. Switzerland



IMDRF

International Medical
Device Regulators Forum

The new EU Regulations on medical devices (adopted 5 April 2017 and published 5 May 2017)

Directive 90/385/EEC on active implantable medical devices

Directive 93/42/EEC on medical devices

Regulation on medical devices (MDR)

Directive 98/79/EC on *in vitro* diagnostic medical devices

Regulation on *in vitro* diagnostic medical devices (IVDR)



Main novelties of the new Regulations (1)

- Inclusion of certain aesthetic devices within the scope.
- EU minimum requirements related to reprocessing of single-use devices.
- Stricter pre-market control of high-risk devices with the involvement of a pool of experts at EU level.
- Reinforcement of the rules on clinical evaluation (and performance evaluation) and clinical investigation (and performance studies).
- Stricter requirements on the use of hazardous substances for certain devices.



Main novelties of the new Regulations (2)

- New classification system for IVDs based on international guidance (80% of IVDs to be assessed by a Notified Body).
- Reinforced designation and oversight processes of notified bodies.
- Clarification of the role and responsibilities of economic operators.
- Establishment of a comprehensive EU database on medical devices (EUDAMED) with large part of information to be made publicly available.
- Introduction of a UDI system.
- Enhanced cooperation amongst national authorities.
- Stronger coordination role of the European Commission.




IMDRF

International Medical
Device Regulators Forum

Towards implementation



Transitional period





COM implementation priorities (1)

- **Notified Bodies**
 - ✓ Launch of designation procedure (November 2017)
 - ✓ 42 applications received up to date. Full scope of MDR and IVR covered
 - ✓ 1st Notified Body designated in Feb 2019
- **Governance**
 - ✓ Setting up of MDCG (November 2017)
 - ✓ MDCG technical subgroups operational as from 1st March 2019
- **Scientific structures**
 - ✓ Establishment of expert panels, expert laboratories and reference labs
- **Design and establishment of the new EUDAMED**
 - ✓ Plan for implementation of functional specifications (May 2018)
 - ✓ Functional specifications (work ongoing)
- **Establishment of UDI system**
 - ✓ First guidelines published, nomenclature selected in Feb 2019, designation of issuing entities to be completed by May 2019



COM implementation priorities (2)

- **Mandate for revision of standards** (Q2 2019)
- **Communication campaign**
 - ✓ The new dedicated website and first updated library are live
 - ✓ Release of existing factsheets in some major non-EU languages has also started.
- **Common specifications on devices without medical purpose** (expected Q1 2020)
- **Common specifications on reprocessing of single-use devices** (November 2019)

Planning of activities:

- **Publication of Commission's rolling plan on DG GROW website**



IMDRF

International Medical
Device Regulators Forum

Useful links

ec.europa.eu

> growth > sectors

> register of Commission expert groups >
mdcg

> law > better-regulation > have-your-say

camd-europe.eu

> MDR/IVDR implementation



IMDRF

International Medical
Device Regulators Forum

Thank you for your attention !

Erik Hansson

European Commission
Health Technology and Cosmetics